



CERTIFICATE

< Member No. 4120 >

Company: TQ-Systems GmbH

Location of Facility:

Buergermeister-Ulrich-Str.100, 86199 Augsburg, Germany

Registration No.: C-12053

Facility Name: Product Compliance Center / Shielded Room No. 2

(Facility for measuring conducted emissions at AC mains power ports)

Registered Instruments: Vertical Reference Plane

This is to certify that the above measuring facility is registered in accordance with VCCI Council Rules.

*** comments on appendix attached**

Date of Registration: June 24, 2022

This Certificate is valid until June 23, 2025

VCCI Council
Akira Oda

Akira Oda





TQ-Systems GmbH (Member No. 4120)

To: Site Manager

Facility Name: Product Compliance Center / Shielded Room No. 2

REGISTRATION No.: C-12053

Comments on Registration

We hereby notify that the Measurement Facility shown in the Certificate is fully examined and approved by the Registration Committee for Measurement Facilities. However, the Committee made the following comment(s) to take into consideration.

We recommend that you calibrate active (electronic) measurement equipment at least once a year.

We recommend that you calibrate passive (non-electronic) measurement equipment at least once every two years.



CERTIFICATE

< Member No. 4120 >

Company: TQ-Systems GmbH

Location of Facility:

Buergermeister-Ulrich-Strasse 100, 86199 Augsburg, Germany

Registration No.: G-20082

Facility Name: FSOATS

(Facility for measuring radiated emissions above 1 GHz)

Frequency Range: 1 GHz - 6 GHz

Measurement Distance: 3 m

Test Volume Diameter: 2.0 m

This is to certify that the above measuring facility is registered in accordance with VCCI Council Rules.

*** comments on appendix attached**

Date of Registration: November 18, 2022

This Certificate is valid until November 17, 2025

VCCI Council
Akira Oda

Akira Oda





TQ-Systems GmbH (Member No. 4120)

To: Site Manager

Facility Name: FSOATS

REGISTRATION No.: G-20082

Comments on Registration

We hereby notify that the Measurement Facility shown in the Certificate is fully examined and approved by the Registration Committee for Measurement Facilities. However, the Committee made the following comment(s) to take into consideration.

We recommend that you calibrate active (electronic) measurement equipment at least once a year.

We recommend that you calibrate passive (non-electronic) measurement equipment at least once every two years.

We recommend conducting test site assessment within one year of calibrating the active (electronic) measurement equipment to be used for assessment.



CERTIFICATE

< Member No. 4120 >

Company: TQ-Systems GmbH

Location of Facility:

Buergermeister-Ulrich-Strasse 100, 86199 Ausburg, Germany

Registration No.: R-20082

Facility Name: SAC

(Facility for measuring radiated emissions below 1 GHz)

Measurement Distance: 10 m

Type of the facility: Semi Anechoic Chamber

This is to certify that the above measuring facility is registered in accordance with VCCI Council Rules.

*** comments on appendix attached**

Date of Registration: September 2, 2022

This Certificate is valid until September 1, 2025

VCCI Council
Akira Oda

Akira Oda





TQ-Systems GmbH (Member No. 4120)

To: Site Manager

Facility Name: SAC

REGISTRATION No.: R-20082

Comments on Registration

We hereby notify that the Measurement Facility shown in the Certificate is fully examined and approved by the Registration Committee for Measurement Facilities. However, the Committee made the following comment(s) to take into consideration.

We recommend that you calibrate active (electronic) measurement equipment at least once a year.

We recommend that you calibrate passive (non-electronic) measurement equipment at least once every two years.



CERTIFICATE

< Member No. 4120 >

Company: TQ-Systems GmbH

Location of Facility:

Buergermeister-Ulrich-Str. 100, 86199 Augsburg, Germany

Registration No.: T-20059

Facility Name: Shielded Room No. 2

(Facility for measuring conducted emissions at Telecommunication (Wired) ports)

Registered Instruments: Vertical Reference Plane

AAN[Transformer type 4 pairs 75 dB, Screened type]

This is to certify that the above measuring facility is registered in accordance with VCCI Council Rules.

*** comments on appendix attached**

Date of Registration: September 2, 2022

This Certificate is valid until September 1, 2025

VCCI Council

Akira Oda

Akira Oda





TQ-Systems GmbH (Member No. 4120) To: Site

Manager

Facility Name: Shielded Room No.2

REGISTRATION No.: T-20059

Comments on Registration

We hereby notify that the Measurement Facility shown in the Certificate is fully examined and approved by the Registration Committee for Measurement Facilities. However, the Committee made the following comment(s) to take into consideration.

We recommend that you calibrate active (electronic) measurement equipment at least once a year.

We recommend that you calibrate passive (non-electronic) measurement equipment at least once every two years.